

Notice of Allowability

Application No.

09/930,582

Examiner

Michael G. Mendoza

Applicant(s)

DENYER ET AL.

Art Unit

3734

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to amendment filed 10-31-2007.
2. ☒ The allowed claim(s) is/are 9-11 and 15-17.
3. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☒ All b) ☐ Some* c) ☐ None of the:
 1. ☒ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 5. ☒ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☒ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☐ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☒ Information Disclosure Statements (PTO/SB/08),
Paper No./Mail Date 10/31/07, 1126/07
4. ☐ Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413),
Paper No./Mail Date _____.
7. ☒ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____.

REASONS FOR ALLOWANCE

1. The following is an examiner's statement of reasons for allowance: the prior art of record fails to teach or render obvious the overall claimed invention of a system or method for delivery of a drug to a patient comprising: a trend generator for analyzing the breath information and the characteristics of the patient's breathing over a number of treatment, wherein the breath information includes a total treatment time, a total inhalation time and a total exhalation time, and wherein a data analyzer calculates a total no flow time value by subtracting the total inhalation time and the total exhalation time from the total treatment time along with other limitations of the claim.
2. The prior art of record fails to teach or render obvious the overall claimed invention of a system or method for delivery of a drug to a patient comprising: a trend generator for analyzing the breath information and the characteristics of the patient's breathing over a number of treatment, wherein the breath information includes a number of breaths per minute and a mean tidal volume, and wherein a data analyzer calculates a minute volume value by dividing the number of breathes per minute by the mean tidal volume along with other limitations of the claim.
3. The prior art of record fails to teach or render obvious the overall claimed invention of a system or method for delivery of a drug to a patient comprising: a trend generator for analyzing the breath information and the characteristics of the patient's breathing over a number of treatment, wherein the breath information includes a number of treatment, wherein the breath information further includes a mean inhalation flow and a mean inhalation time, and wherein a data analyzer calculates a mean tidal

volume value by multiplying the mean inhalation flow with the mean inhalation time with 0.7 and dividing by 60.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Drawings

New corrected drawings in compliance with 37 CFR 1.121(d) are required in this application because of handwritten text and numbers. Applicant is advised to employ the services of a competent patent draftsman outside the Office, as the U.S. Patent and Trademark Office no longer prepares new drawings. The corrected drawings are required in reply to the Office action to avoid abandonment of the application. The requirement for corrected drawings will not be held in abeyance.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael G. Mendoza whose telephone number is (571) 272-4698. The examiner can normally be reached on Mon.-Fri. 9:00 a.m. - 5:00 p.m..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Michael Hayes can be reached on (571) 272-4959. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number:
09/930,582
Art Unit: 3734

Page 4

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

mm

MM

M J Hayes

MICHAEL J. HAYES
SUPERVISORY PATENT EXAMINER